

NI Serial Hardware Specifications Guide

This document lists safety and compliance information for NI Serial hardware, as well as physical specifications, software characteristics, and recommended operating conditions.

Contents

Safety and Radio-Frequency Interference.....	1
PXI Serial Hardware	9
USB Serial Hardware.....	14
ENET Serial Hardware	17
PCMCIA Serial Hardware	19
AT Serial Hardware	21
Technical Support Resources.....	25

Safety and Radio-Frequency Interference

This section contains safety instructions and information about the radio-frequency interference characteristics of the hardware it accompanies. Read this section before installing and using the new hardware.

Safety Information

This section contains important safety information that you *must* follow when installing and using the module.

Do *not* operate the module in a manner not specified in this document. Misuse of the module can result in a hazard. You can compromise the safety protection built into the module if the module is damaged in any way. If the module is damaged, return it to National Instruments (NI) for repair.

Do *not* substitute parts or modify the module except as described in this document. Use the module only with the chassis, modules, accessories, and cables specified in the installation instructions. You *must* have all covers and filler panels installed during operation of the module.

Do *not* operate the module in an explosive atmosphere or where there may be flammable gases or fumes. If you must operate the module in such an environment, it must be in a suitably rated enclosure.

If you need to clean the module, use a soft, nonmetallic brush. Make sure that the module is completely dry and free from contaminants before returning it to service.

Operate the module only at or below Pollution Degree 2. Pollution is foreign matter in a solid, liquid, or gaseous state that can reduce dielectric strength or surface resistivity. The following is a description of pollution degrees:

- Pollution Degree 1 means no pollution or only dry, nonconductive pollution occurs. The pollution has no influence.
- Pollution Degree 2 means that only nonconductive pollution occurs in most cases. Occasionally, however, a temporary conductivity caused by condensation must be expected.
- Pollution Degree 3 means that conductive pollution occurs, or dry, nonconductive pollution occurs that becomes conductive due to condensation.

You *must* insulate signal connections for the maximum voltage for which the module is rated. Do *not* exceed the maximum ratings for the module. Do not install wiring while the module is live with electrical signals. Do not remove or add connector blocks when power is connected to the system. Avoid contact between your body and the connector block signal when hot swapping modules. Remove power from signal lines before connecting them to or disconnecting them from the module.

FCC/Canada Radio Frequency Interference Compliance

Determining FCC Class

The Federal Communications Commission (FCC) has rules to protect wireless communications from interference. The FCC places digital electronics into two classes. These classes are known as Class A (for use in industrial-commercial locations only) or Class B (for use in residential or commercial locations). All National Instruments (NI) products are FCC Class A products.

Depending on where it is operated, this Class A product could be subject to restrictions in the FCC rules. (In Canada, the Department of Communications (DOC), of Industry Canada, regulates wireless interference in much the same way.) Digital electronics emit weak signals during normal operation that can affect radio, television, or other wireless products.

All Class A products display a simple warning statement of one paragraph in length regarding interference and undesired operation. The FCC rules have restrictions regarding the locations where FCC Class A products can be operated.

Consult the FCC Web site at www.fcc.gov for more information.

FCC/DOC Warnings

This equipment generates and uses radio frequency energy and, if not installed and used in strict accordance with the instructions in this manual and the CE marking Declaration of Conformity¹, may cause interference to radio and television reception. Classification requirements are the same for the Federal Communications Commission (FCC) and the Canadian Department of Communications (DOC).

Changes or modifications not expressly approved by NI could void the user's authority to operate the equipment under the FCC Rules.

Class A

Federal Communications Commission

This equipment has been tested and found to comply with the limits for a Class A digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference when the equipment is operated in a commercial environment. This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used in accordance with the instruction manual, may cause harmful interference to radio communications. Operation of this equipment in a residential area is likely to cause harmful interference in which case the user is required to correct the interference at their own expense.

¹ The CE marking Declaration of Conformity contains important supplementary information and instructions for the user or installer.

Canadian Department of Communications

This Class A digital apparatus meets all requirements of the Canadian Interference-Causing Equipment Regulations.

Cet appareil numérique de la classe A respecte toutes les exigences du Règlement sur le matériel brouilleur du Canada.

Compliance to EU Directives

Users in the European Union (EU) should refer to the Declaration of Conformity (DoC) for information¹ pertaining to the CE marking. Refer to the Declaration of Conformity (DoC) for this product for any additional regulatory compliance information. To obtain the DoC for this product, visit ni.com/hardref.nsf, search by model number or product line, and click the appropriate link in the Certification column.



Caution If the NI Serial hardware is used in a manner inconsistent with the instructions or specifications listed by National Instruments, the protective features of the chassis may be impaired.

PCI Serial Hardware

This section describes the characteristics of the PCI serial hardware and the recommended operating conditions.



Note This equipment is intended for indoor use only.

Safety

The NI Serial hardware has been evaluated using the criteria of EN 61010-1 and meets the requirements of the following standards for safety and electrical equipment for measurement, control, and laboratory use:

- IEC 60950-1, EN 60950-1
- UL 60950-1
- CAN/CSA-C22.2 No. 60950-1



Note For UL and other safety certifications, refer to the product label, or visit ni.com/certification, search by model number or product line, and click the appropriate link in the Certification column.

¹ The CE marking Declaration of Conformity contains important supplementary information and instructions for the user or installer.

Electromagnetic Compatibility

Emissions	EN 55011 Class A at 10 m FCC Part 15A above 1 GHz
Immunity	EN 61326:1997 + A2:2001, Table 1
EMC/EMI.....	CE, C-Tick, and FCC Part 15 (Class A) Compliant



Note For EMC compliance, operate this device with shielded cabling. In addition, all covers and filler panels must be installed.

CE Compliance

This product meets the essential requirements of applicable European Directives, as amended for CE marking, as follows:

Low-Voltage Directive (safety) 73/23/EEC

Electromagnetic Compatibility
Directive (EMC) 89/336/EEC



Note Refer to the Declaration of Conformity (DoC) for this product for any additional regulatory compliance information. To obtain the DoC for this product, visit ni.com/certification, search by model number or product line, and click the appropriate link in the Certification column.

Nonisolated PCI Two-Port Boards

Dimensions..... 10.67 by 14.22 cm
(4.2 by 5.6 in.)

I/O connector..... DB-9

Power requirement (from PCI channel)

PCI-485/2

+5 VDC..... 350 mA typical
750 mA maximum

PCI-232/2

+5 VDC..... 50 mA typical
100 mA maximum

±12 VDC..... 20 mA typical
200 mA maximum

PCI-8430/2	
+5 VDC	325 mA typical 500 mA maximum
PCI-8431/2	
+5 VDC	500 mA typical 700 mA maximum

Nonisolated PCI Four-Port Boards

Dimensions	10.67 by 17.27 cm (4.2 by 5.6 in.)
I/O connector ¹	10-position modular jack
Power requirement (from PCI channel)	
PCI-485/4	
+5 VDC	700 mA typical 1,300 mA maximum
PCI-232/4	
+5 VDC	70 mA typical 150 mA maximum
±12 VDC	40 mA typical 400 mA maximum
PCI-8430/4	
+5 VDC	400 mA typical 600 mA maximum
PCI-8431/4	
+5 VDC	725 mA typical 1.1 A maximum

Nonisolated PCI Eight-Port Boards

Dimensions	10.67 by 14.48 cm (4.2 by 5.7 in.)
I/O connector ²	68-position, SCSI type connector

¹ The four-port PCI serial boards require a cable to convert the 10-position modular jack to either DB-9 or DB-25 connectors.

² The eight-port PCI serial boards require a cable, which is included in your kit, to convert the 68-position connector to eight DB-9 connectors.

Power requirement (from PCI channel)

PCI-485/8

+5 VDC..... 1,100 mA typical
2,000 mA maximum

PCI-232/8

+5 VDC..... 100 mA typical
180 mA maximum

±12 VDC..... 80 mA typical
800 mA maximum

PCI-8430/8

+5 VDC..... 600 mA typical
900 mA maximum

PCI-8431/8

+5 VDC..... 1.3 mA typical
1.9 A maximum

Nonisolated PCI 16-Port Boards

Dimensions..... 10.67 by 17.52 cm
(4.2 by 6.9 in.)

I/O connector¹ 100-position, SCSI type
connector

Power requirement (from PCI channel)

PCI-232/16

+5 VDC..... 250 mA typical
500 mA maximum

Isolated PCI Two-Port Boards

Dimensions..... 10.67 by 14.22 cm
(4.2 by 6.9 in.)

I/O connector..... DB-9

Isolation voltage

From port to port..... 2,000 V_{rms}/60 s

From any port to host computer..... 2,000 V_{rms}/60 s

¹ The 16-port PCI serial boards require a breakout box, which is included in your kit, to separate the 100-position connector to 16 DB-9 connectors.

Power requirement (from PCI channel)

PCI-485/2

+5 VDC800 mA typical
1,300 mA maximum

PCI-232/2

+5 VDC400 mA typical
650 mA maximum

Isolated PCI Four-Port Boards

Dimensions10.67 by 17.27 cm
(4.2 by 6.9 in.)

I/O connector¹10-position modular jack

Isolation voltage

From port to port2,000 V_{rms}/60 s

From any port to host computer2,000 V_{rms}/60 s

Power requirement (from PCI channel)

PCI-485/4

+5 VDC1,000 mA typical
1,500 mA maximum

PCI-232/4

+5 VDC500 mA typical
750 mA maximum

Environmental Characteristics

Operating Environment

Ambient temperature0 to 55 °C (Tested in accordance
with IEC-60068-2-1 and
IEC-60068-2-2.)

Relative humidity10 to 90%, noncondensing
(Tested in accordance with
IEC-60068-2-56.)

Altitude (maximum)2,000 m

¹ The four-port PCI serial boards require a cable to convert the 10-position modular jack to either DB-9 or DB-25 connectors.

Storage Environment

Ambient temperature.....	-20 to 70 °C (Tested in accordance with IEC-60068-2-1 and IEC-60068-2-2.)
Relative humidity.....	5 to 95%, noncondensing (Tested in accordance with IEC-60068-2-56.)

PXI Serial Hardware

This section describes the characteristics of the PXI serial hardware and the recommended operating conditions.



Note This equipment is intended for indoor use only.

Safety

The NI Serial hardware has been evaluated using the criteria of EN 61010-1 and meets the requirements of the following standards for safety and electrical equipment for measurement, control, and laboratory use:

- IEC 60950-1, EN 60950-1
- UL 60950-1
- CAN/CSA-C22.2 No. 60950-1



Note For UL and other safety certifications, refer to the product label, or visit ni.com/certification, search by model number or product line, and click the appropriate link in the Certification column.

Electromagnetic Compatibility

Emissions	EN 55011 Class A at 10 m FCC Part 15A above 1 GHz
Immunity.....	EN 61326:1997 + A2:2001, Table 1
EMC/EMI.....	CE, C-Tick, and FCC Part 15 (Class A) Compliant



Note For EMC compliance, operate this device with shielded cabling. In addition, all covers and filler panels must be installed.

CE Compliance

This product meets the essential requirements of applicable European Directives, as amended for CE marking, as follows:

Low-Voltage Directive (safety).....73/23/EEC

Electromagnetic Compatibility
Directive (EMC)89/336/EEC



Note Refer to the Declaration of Conformity (DoC) for this product for any additional regulatory compliance information. To obtain the DoC for this product, visit ni.com/certification, search by model number or product line, and click the appropriate link in the Certification column.

Nonisolated PXI Two-Port Boards

Dimensions100 by 160 mm
(3.94 by 6.37 in.)

I/O connectorDB-9

Power requirement (from PXI channel)

PXI-8420/2

+5 VDC100 mA typical
150 mA maximum

±12 VDC20 mA typical
200 mA maximum

PXI-8421/2

+5 VDC350 mA typical
750 mA maximum

PXI-8430/2

+5 VDC325 mA typical
500 mA maximum

PXI-8431/2

+5 VDC500 mA typical
750 mA maximum

Nonisolated PXI Four-Port Boards

Dimensions.....	100 by 160 mm (3.94 by 6.37 in.)
I/O connector ¹	10-position modular jack
Power requirement (from PXI channel)	
PXI-8420/4	
+5 VDC.....	125 mA typical 200 mA maximum
±12 VDC.....	40 mA typical 400 mA maximum
PXI-8421/4	
+5 VDC.....	350 mA typical 750 mA maximum
PXI-8430/4	
+5 VDC.....	400 mA typical 600 mA maximum
PXI-8431/4	
+5 VDC.....	725 mA typical 1.1 A maximum

Nonisolated PXI Eight-Port Boards

Dimensions.....	100 by 160 mm (3.94 by 6.37 in.)
I/O connector ²	68-position, SCSI type connector
Power requirement (from PXI channel)	
PXI-8420/8	
+5 VDC.....	150 mA typical 250 mA maximum
±12 VDC.....	80 mA typical 800 mA maximum
PXI-8421/8	
+5 VDC.....	1,100 mA typical 2,000 mA maximum

¹ The four-port PXI serial boards require a cable to convert the 10-position modular jack to either DB-9 or DB-25 connectors.

² The eight-port PXI serial boards require cables, which are included in your kit, to convert the 68-position connector to eight DB-9 connectors.

Nonisolated PXI 16-Port Boards

Dimensions	100 by 160 mm (3.94 by 6.37 in.)
I/O connector ¹	100-position, SCSI type connector
Power requirement (from PXI channel)	
PXI-8420/16	
+5 VDC	500 mA typical 750 mA maximum

Isolated PXI Two-Port Boards

Dimensions	100 by 160 mm (3.94 by 6.37 in.)
I/O connector	DB-9
Power requirement (from PXI channel)	
PXI-8422/2	
+5 VDC	400 mA typical 650 mA maximum
PXI-8423/2	
+5 VDC	800 mA typical 1,300 mA maximum
Isolation voltage	
From port to port	2,000 V _{rms} /60 s
From any port to host computer	2,000 V _{rms} /60 s

Isolated PXI Four-Port Boards

Dimensions	100 by 160 mm (3.94 by 6.37 in.)
I/O connector ²	10-position modular jack

¹ The 16-port PCI serial boards require a breakout box, which is included in your kit, to separate the 100-position connector to 16 DB-9 connectors.

² The four-port PXI serial boards require a cable to convert the 10-position modular jack to either DB-9 or DB-25 connectors.

Power requirement (from PXI channel)

PXI-8422/4

+5 VDC..... 500 mA typical
750 mA maximum

PXI-8423/4

+5 VDC..... 1,000 mA typical
1,500 mA maximum

Isolation voltage

From port to port..... 2,000 V_{rms}/60 s

From any port to host computer..... 2,000 V_{rms}/60 s

Environmental Characteristics

Operating Environment

Ambient temperature..... 0 to 55 °C (Tested in accordance with IEC-60068-2-1 and IEC-60068-2-2.)

Relative humidity 10 to 90%, noncondensing (Tested in accordance with IEC-60068-2-56.)

Altitude (maximum)..... 2,000 m

Storage Environment

Ambient temperature..... -20 to 70 °C (Tested in accordance with IEC-60068-2-1 and IEC-60068-2-2.)

Relative humidity 5 to 95%, noncondensing (Tested in accordance with IEC-60068-2-56.)

USB Serial Hardware

This section describes the characteristics of the USB serial hardware and the recommended operating conditions.



Note This equipment is intended for indoor use only.

Safety

The NI Serial hardware has been evaluated using the criteria of EN 61010-1 and meets the requirements of the following standards for safety and electrical equipment for measurement, control, and laboratory use:

- IEC 60950-1, EN 60950-1
- UL 60950-1
- CAN/CSA-C22.2 No. 60950-1



Note For UL and other safety certifications, refer to the product label, or visit ni.com/certification, search by model number or product line, and click the appropriate link in the Certification column.

Electromagnetic Compatibility

Emissions	EN 55011 Class A at 10 m FCC Part 15A above 1 GHz
Immunity	EN 61326:1997 + A2:2001, Table 1
EMC/EMI	CE, C-Tick, and FCC Part 15 (Class A) Compliant



Note For EMC compliance, operate this device with shielded cabling. In addition, all covers and filler panels must be installed.

CE Compliance

This product meets the essential requirements of applicable European Directives, as amended for CE marking, as follows:

Low-Voltage Directive (safety)	73/23/EEC
Electromagnetic Compatibility Directive (EMC)	89/336/EEC



Note Refer to the Declaration of Conformity (DoC) for this product for any additional regulatory compliance information. To obtain the DoC for this product, visit ni.com/certification, search by model number or product line, and click the appropriate link in the Certification column.

One-Port USB Hardware

Dimensions.....	3.8 by 3.7 by 1.6 cm (1.49 by 1.44 by 0.61 in.)
Case material.....	PVC
Weight	
USB-232	121 g (0.27 lb)
USB-485	118 g (0.26 lb)
I/O connector.....	DB-9
USB connector	Captive cable with USB series A plug
Power requirement (from USB channel)	
USB-485	
+5 VDC.....	175 mA typical 500 mA maximum
USB-232	
+5 VDC.....	100 mA typical 200 mA maximum

USB Two- and Four-Port Hardware

Dimensions.....	21.0 by 12.4 by 3.7 cm (8.25 by 4.89 by 1.44 in.)
Case material.....	Hard plastic with metal baseplate
Weight.....	375 g (0.83 lb)
I/O connector.....	DB-9
USB connector	USB series B
Power requirement (from USB channel)	
USB-485/2	
+5 VDC.....	300 mA typical 500 mA maximum

USB-232/2	
+5 VDC	200 mA typical 500 mA maximum
USB-232/4	
+5 VDC	300 mA typical 500 mA maximum
Power requirement (from external supply)	
USB-485/4 (9 V–30 V)	
+12 VDC (typical).....	225 mA typical 500 mA maximum

Environmental Characteristics

Operating Environment

Ambient temperature	0 to 70 °C (Tested in accordance with IEC-60068-2-1 and IEC-60068-2-2.)
Relative humidity	10 to 90%, noncondensing (Tested in accordance with IEC-60068-2-56.)
Altitude (maximum)	2,000 m

Storage Environment

Ambient temperature	
One port.....	–40 to 80 °C (Tested in accordance with IEC-60068-2-1 and IEC-60068-2-2.)
Two and four port.....	–40 to 85 °C (Tested in accordance with IEC-60068-2-1 and IEC-60068-2-2.)
Relative humidity	5 to 95%, noncondensing (Tested in accordance with IEC-60068-2-56.)

ENET Serial Hardware

This section describes the characteristics of the USB serial hardware, along with the recommended operating conditions.



Note This equipment is intended for indoor use only.

Safety

The NI Serial hardware has been evaluated using the criteria of EN 61010-1 and meets the requirements of the following standards for safety and electrical equipment for measurement, control, and laboratory use:

- IEC 60950-1, EN 60950-1
- UL 60950-1
- CAN/CSA-C22.2 No. 60950-1



Note For UL and other safety certifications, refer to the product label, or visit ni.com/certification, search by model number or product line, and click the appropriate link in the Certification column.

Electromagnetic Compatibility

Emissions	EN 55011 Class A at 10 m FCC Part 15A above 1 GHz
Immunity	EN 61326:1997 + A2:2001, Table 1
EMC/EMI.....	CE, C-Tick, and FCC Part 15 (Class A) Compliant



Note For EMC compliance, operate this device with shielded cabling. In addition, all covers and filler panels must be installed.

CE Compliance

This product meets the essential requirements of applicable European Directives, as amended for CE marking, as follows:

Low-Voltage Directive (safety)	73/23/EEC
Electromagnetic Compatibility Directive (EMC)	89/336/EEC



Note Refer to the Declaration of Conformity (DoC) for this product for any additional regulatory compliance information. To obtain the DoC for this product, visit ni.com/certification, search by model number or product line, and click the appropriate link in the Certification column.

Electrical Characteristics

Power requirement (from external supply)	
External supply (9 V–30 V)	
+12 VDC (typical).....	500 mA typical 750 mA maximum

Environmental Characteristics

Operating Environment

Ambient temperature	0 to 70 °C (Tested in accordance with IEC-60068-2-1 and IEC-60068-2-2.)
Relative humidity	10 to 90%, noncondensing (Tested in accordance with IEC-60068-2-56.)
Altitude (maximum)	2,000 m

Storage Environment

Ambient temperature	–40 to 85 °C (Tested in accordance with IEC-60068-2-1 and IEC-60068-2-2.)
Relative humidity	5 to 95%, noncondensing (Tested in accordance with IEC-60068-2-56.)

Physical Characteristics

Overall case size (dimensions)	21.0 by 12.4 by 3.7 cm (8.25 by 4.89 by 1.44 in.)
Case material	Hard plastic with metal baseplate
Weight	394 g (0.87 lb)
Serial connectors.....	DB-9

Network Specifications

Ethernet connector	RJ-45
Connection type	IEEE 802.3 compliant 100Base-TX (100 Mbits/s) 10Base-T (10 Mbits/s)
Duplex mode	Half duplex

PCMCIA Serial Hardware

This section describes the characteristics of the PCMCIA serial hardware, along with the recommended operating conditions.



Note This equipment is intended for indoor use only.

Safety

The NI Serial hardware has been evaluated using the criteria of EN 61010-1 and meets the requirements of the following standards for safety and electrical equipment for measurement, control, and laboratory use:

- IEC 60950-1, EN 60950-1
- UL 60950-1
- CAN/CSA-C22.2 No. 60950-1



Note For UL and other safety certifications, refer to the product label, or visit ni.com/certification, search by model number or product line, and click the appropriate link in the Certification column.

Electromagnetic Compatibility

Emissions	EN 55011 Class A at 10 m FCC Part 15A above 1 GHz
Immunity	EN 61326:1997 + A2:2001, Table 1
EMC/EMI	CE, C-Tick, and FCC Part 15 (Class A) Compliant



Note For EMC compliance, operate this device with shielded cabling. In addition, all covers and filler panels must be installed.

CE Compliance

This product meets the essential requirements of applicable European Directives, as amended for CE marking, as follows:

Low-Voltage Directive (safety).....73/23/EEC

Electromagnetic Compatibility
Directive (EMC)89/336/EEC



Note Refer to the Declaration of Conformity (DoC) for this product for any additional regulatory compliance information. To obtain the DoC for this product, visit ni.com/certification, search by model number or product line, and click the appropriate link in the Certification column.

Hardware Specifications

Dimensions	Type II PC card
I/O connector	Adapter cable with DB-9 Dsub connector and converter for PC card
Power requirement (from PCMCIA expansion slot)	
PCMCIA-232	
+5 VDC	40 mA typical, 150 mA maximum
PCMCIA-485	
+5 VDC	110 mA typical, 225 mA maximum
PCMCIA-232/2	
+5 VDC	60 mA typical, 250 mA maximum
PCMCIA-485/2	
+5 VDC	150 mA typical, 400 mA maximum
PCMCIA-232/4	
+5 VDC	60 mA typical, 200 mA maximum

Environmental Characteristics

Operating Environment

Ambient temperature.....	0 to 55 °C (Tested in accordance with IEC-60068-2-1 and IEC-60068-2-2.)
Relative humidity	10 to 90%, noncondensing (Tested in accordance with IEC-60068-2-56.)
Altitude (maximum).....	2,000 m

Storage Environment

Ambient temperature.....	–40 to 120 °C (Tested in accordance with IEC-60068-2-1 and IEC-60068-2-2.)
Relative humidity	5 to 95%, noncondensing (Tested in accordance with IEC-60068-2-56.)

AT Serial Hardware

This section describes the characteristics of the AT (ISA) serial hardware, along with the recommended operating conditions.



Note This equipment is intended for indoor use only.

Safety

The NI Serial hardware has been evaluated using the criteria of EN 61010-1 and meets the requirements of the following standards for safety and electrical equipment for measurement, control, and laboratory use:

- IEC 60950-1, EN 60950-1
- UL 60950-1
- CAN/CSA-C22.2 No. 60950-1



Note For UL and other safety certifications, refer to the product label, or visit ni.com/certification, search by model number or product line, and click the appropriate link in the Certification column.

Electromagnetic Compatibility

Emissions	EN 55011 Class A at 10 m FCC Part 15A above 1 GHz
Immunity	EN 61326:1997 + A2:2001, Table 1
EMC/EMI	CE, C-Tick, and FCC Part 15 (Class A) Compliant



Note For EMC compliance, operate this device with shielded cabling. In addition, all covers and filler panels must be installed.

CE Compliance

This product meets the essential requirements of applicable European Directives, as amended for CE marking, as follows:

Low-Voltage Directive (safety).....73/23/EEC

Electromagnetic Compatibility
Directive (EMC).....89/336/EEC



Note Refer to the Declaration of Conformity (DoC) for this product for any additional regulatory compliance information. To obtain the DoC for this product, visit ni.com/certification, search by model number or product line, and click the appropriate link in the Certification column.

Nonisolated ISA Two-Port Boards

Dimensions

10.67 by 16.51 cm (4.2 by 6.5 in.)

I/O connector

DB-9

Power requirement
(from PC AT I/O channel)

AT-485/2

+5 VDC	390 mA typical 510 mA maximum
--------------	----------------------------------

AT-232/2

+5 VDC	260 mA typical 340 mA maximum
--------------	----------------------------------

AT-485/2 (Shared IRQ)	
+5 VDC.....	140 mA typical 180 mA maximum
AT-232/2 (Shared IRQ)	
+5 VDC.....	70 mA typical 100 mA maximum

Nonisolated ISA Four-Port Boards

Dimensions.....	10.67 by 16.51 cm (4.2 by 6.5 in.)
I/O connector ¹	10-position modular jack
Power requirement (from PC AT I/O channel)	
AT-485/4	
+5 VDC.....	160 mA typical 200 mA maximum
AT-232/4	
+5 VDC.....	110 mA typical 150 mA maximum
AT-485/4 (Shared IRQ)	
+5 VDC.....	160 mA typical 200 mA maximum
AT-232/4 (Shared IRQ)	
+5 VDC.....	110 mA typical 150 mA maximum

Isolated ISA Two-Port Boards

Dimensions.....	10.67 by 18.70 cm (4.2 by 6.9 in.)
I/O connector.....	DB-9
Isolation voltage	
From Port to Port	2,000 V _{rms} /60 s
From Any Port to Host Computer	330 V _{rms} /60 s

¹ The four-port AT serial board requires a cable to convert the 10-position modular jack to either DB-9 or DB-25 connectors.

Power requirement
(from PC AT I/O channel)

AT-485/2 Isolated

+5 VDC220 mA typical
260 mA maximum

AT-232/2 Isolated

+5 VDC160 mA typical
200 mA maximum

Isolated Four-Port Boards

Dimensions10.67 by 25.40 cm
(4.2 by 10.0 in.)

I/O connector¹10-position modular jack

Isolation voltage

From Port to Port2,000 V_{rms}/60 s

From Any Port to Host

Computer2000 V_{rms}/60 s

Power requirement
(from PC AT I/O channel)

AT-485/4 Isolated

+5 VDC300 mA typical
360 mA maximum

AT-232/4 Isolated

+5 VDC280 mA typical
320 mA maximum

Environmental Characteristics

Operating Environment

Ambient temperature0 to 55 °C (Tested in accordance
with IEC-60068-2-1 and
IEC-60068-2-2.)

Relative humidity10 to 90%, noncondensing
(Tested in accordance with
IEC-60068-2-56.)

Altitude (maximum)2,000 m

¹ The four-port isolated AT serial board requires cables, which are included in your kit, to convert the 10-position modular jack to a DB-9 connector. Use only the type of cables provided in your kit.

Storage Environment

Ambient temperature.....	–40 to 120 °C (Tested in accordance with IEC-60068-2-1 and IEC-60068-2-2.)
Relative humidity.....	5 to 95%, noncondensing (Tested in accordance with IEC-60068-2-56.)

Technical Support Resources

NI Web Support

National Instruments Web support is your first stop for help in solving installation, configuration, and application problems and questions. Online problem-solving and diagnostic resources include frequently asked questions, knowledge bases, product-specific troubleshooting wizards, manuals, drivers, software updates, and more. Web support is available through the Technical Support section of ni.com.

A Declaration of Conformity (DoC) is our claim of compliance with the Council of the European Communities using the manufacturer's declaration of conformity. This system affords the user protection for electronic compatibility (EMC) and product safety. You can obtain the DoC for your product by visiting ni.com/certification. If your product supports calibration, you can obtain the calibration certificate for your product at ni.com/calibration.

Worldwide Support

National Instruments corporate headquarters is located at 11500 North Mopac Expressway, Austin, Texas, 78759-3504. National Instruments also has offices located around the world to help address your support needs. You can access our branch office Web sites from the Worldwide Offices section of ni.com. Branch office Web sites provide up-to-date contact information, support phone numbers, email addresses, and current events.

If you have searched the technical support resources on our Web site and still cannot find the answers you need, contact your local office or National Instruments corporate. For telephone support in the United States, dial 512 795 8248. For telephone support outside the United States, contact your local branch office:

Australia 1800 300 800, Austria 43 0 662 45 79 90 0,
Belgium 32 0 2 757 00 20, Brazil 55 11 3262 3599,

Canada (Calgary) 403 274 9391, Canada (Ottawa) 613 233 5949,
Canada (Québec) 450 510 3055, Canada (Toronto) 905 785 0085,
Canada (Vancouver) 604 685 7530, China 86 21 6555 7838,
Czech Republic 420 224 235 774, Denmark 45 45 76 26 00,
Finland 385 0 9 725 725 11, France 33 0 1 48 14 24 24,
Germany 49 0 89 741 31 30, India 91 80 51190000,
Israel 972 0 3 6393737, Italy 39 02 413091, Japan 81 3 5472 2970,
Korea 82 02 3451 3400, Malaysia 603 9131 0918,
Mexico 01 800 010 0793, Netherlands 31 0 348 433 466,
New Zealand 0800 553 322, Norway 47 0 66 90 76 60,
Poland 48 22 3390150, Portugal 351 210 311 210,
Russia 7 095 783 68 51, Singapore 65 6226 5886,
Slovenia 386 3 425 4200, South Africa 27 0 11 805 8197,
Spain 34 91 640 0085, Sweden 46 0 8 587 895 00,
Switzerland 41 56 200 51 51, Taiwan 886 2 2528 7227,
Thailand 662 992 7519, United Kingdom 44 0 1635 523545

National Instruments™, NI™, and ni.com™ are trademarks of National Instruments Corporation. Product and company names mentioned herein are trademarks or trade names of their respective companies. For patents covering National Instruments products, refer to the appropriate location: **Help»Patents** in your software, the `patents.txt` file on your CD, or ni.com/patents.