

RELEASE NOTES

Calibration Executive

Version 3.3

Known Issues

The following list contains information about known issues with Calibration Executive 3.3.

- If you have NI TestStand Development System 4.0 or 4.0.1 installed, you must install the NI TestStand Development System 4.0.1f1 before installing Calibration Executive 3.3. Refer to the National Instruments Web site at ni.com/support to download the 4.0.1f1 patch.
- If you uninstall NI TestStand Development System 4.0.1f1, you will need to reinstall Calibration Executive 3.3.

Upgrade and Compatibility Issues

You might encounter the following issue if you revert to a previous version of Calibration Executive.

Calibration Reports Database

Previous versions of Calibration Executive do not support the report database file (`calibration_reports.mdb`) that is in version 3.3. If you plan to revert to a previous version of Calibration Executive, refer to the *Calibration Reports Database* topic in the *Calibration Executive Help* for information on saving a database that is compatible with previous versions.

After installing version 3.2 or prior, you can specify this file to retrieve calibration records or add new ones. Refer to the *Configuring Database Options* topic in the *Calibration Executive Help* for information on specifying the correct file.

National Instruments, NI, ni.com, and LabVIEW are trademarks of National Instruments Corporation. Refer to the *Terms of Use* section on ni.com/legal for more information about National Instruments trademarks. Other product and company names mentioned herein are trademarks or trade names of their respective companies. For patents covering National Instruments products, refer to the appropriate location: **Help»Patents** in your software, the `patents.txt` file on your media, or ni.com/patents.